

A large, stylized graphic of a DNA double helix in shades of purple and white, spanning the width of the top half of the page.

TGuide Smart DNA Purification Kit

For recovery of DNA from PCR product or agarose gels.

TECHNICAL MANUAL

Cat. no. 4993550

Note: To use the TGuide Smart DNA purification Kit, you must have the TGuide Smart DNA purification Kit (program no. DP642) installed on the TGuide S16/S32 pro Nucleic Acid Extractor.



www.tiangen.com/en

This product is for scientific research use only.
Do not use in medicine, clinical treatment, food
or cosmetic

Table Contents

<i>Kit Contents</i>	1
<i>DNA purification reagent composition</i>	1
<i>Storage condition</i>	1
<i>Product</i>	2
<i>Features</i>	2
<i>Notes</i>	2
<i>Operational steps</i>	3
<i>1. Prefilled single sample cartridge</i>	4
<i>2. Sample preparation</i>	4
<i>3. Operation steps of TGuide S16 Nucleic Acid Extractor</i>	4
<i>Appendix</i>	5
<i>1. Program</i>	5
<i>2. Related Products</i>	5

TGuide Smart DNA Purification Kit

Cat. no. 4993550

Kit Contents

Contents		4993550 (48 preps)
4993543	DNA Purification Reagents	48
	Buffer TB	15 ml
4993546	TGuide Smart Tip Comb	12 pcs

Note: 4993546 is shipped and packaged separately

DNA purification reagent composition

Well 1	Well 2	Well 3	Well 4	Well 5	Well 6
Buffer PC	Buffer PC reserve well	Buffer PWP	Buffer PWP	None	MagAttract Suspension GSP1
500 µl	500 µl	600 µl	600 µl		260 µl

Storage condition

The kit can be stored under dry conditions at room temperature (15~30°C) for 12 months.

Product

This product adopts magnetic beads and a unique buffer system to recover DNA quickly and efficiently from TAE/TBE gel. The uniquely embedded magnetic beads have a strong affinity for nucleic acid under certain conditions. When the conditions are changed, the magnetic beads can release the absorbed nucleic acid to rapidly separate and purify it. The kit can be used to remove impurities such as protein, other organic compounds, inorganic ions and oligonucleotide primers. It can recover 100 bp~15 kb DNA fragments, with a recovery rate of more than 80%.

The DNA recovered by this kit can be used for a series of common downstream applications, including digestion, PCR, sequencing, library screening, ligating, transformations and other experiments.


Features

- **Fast:** Ultra-pure DNA can be obtained by running TGuide S16 for 40 minutes.
- **Diversity:** Single and double strand DNA fragments and circular plasmid DNA.
- **Efficiency:** Unique magnetic beads and specialized buffer make it possible to recover the largest amount of the target DNA of high purity.

Notes

1. When cutting the gel, UV illumination time should be controlled to avoid damage to DNA.
2. TAE electrophoresis buffer is optimal if the next step is more demanding.
3. The recovery rate is affected by the initial DNA input amount and elution volume. Lower initial DNA input amount and elution volume will lead to lower recovery rate.

Operational steps



1. Prefilled single sample cartridge

1.1 Take out a prefilled single sample cartridge and invert it to re-suspend the magnetic beads; Gently shake to concentrate the reagent and magnetic beads to the bottom of the cartridge. Before use, remove sealing film carefully to avoid liquid spatter or spills.

1.2 Add proper volume (30~100 μ l) of elution buffer TB to the 5th well of the cartridge.

Note: That the less elution volume is used, the higher nucleic acid concentration can be obtained, but the lower total yield will be. So we recommend 60 μ l as the optimal elution volume. The volume could be adjusted if there is any special requirement.

2. Sample preparation

2.1 Gel recovery

Cut about 150 mg single target DNA band from the agarose gel with a clean blade (remove extra gel as much as possible). Proceed to 3.1.

2.2 Purification of PCR product


Dilute the PCR product with ddH₂O(not provided) to 150 μ l. Proceed to 3.1.

Note: If the gel slice is more than 150 mg or the volume of PCR product is more than 150 μ l, add 3 times volume of Buffer PC (from the 2nd well), and the total volume of the solution in the 1st well should not exceed 900 μ l.

3. Operation steps of TGuide S16 Nucleic Acid Extractor

3.1 Add the above sample to the 1st well of the cartridge. Put cartridges on the reagent tank bracket of TGuide S16 Nucleic Acid Extractor.

3.2 Place the reagent tank bracket on the plate base in the TGuide S16 Nucleic Acid Extractor. Insert the Tip Combs into the slots to ensure that they are well connected and firmed.



3.3 If you use the TGuide S16 Nucleic Acid Extractor, select the corresponding program DP642 file on the touch screen, click the icon  in the lower right corner of the screen, or click the "RUN" button at the bottom of the screen to start the experiment.

3.4 At the end of the automated extraction process, take the DNA out of the 5 th well of the cartridge and store it under appropriate conditions. Sample reagent cartridge and tip comb are for single use only.

Appendix

1. Program

The extraction process of S16 provided for DP642 is shown in the following table:

Step	Hole site	Step name	Mix time (min)	Mix speed	Dry time (min)	Volume (μl)	Temp. (°C)	Segments	Every time (sec)	Magnetization time(sec)	Cycle	Magnet speed (mm/s)
1	1	Lysis	15	3	0	650	60	1	0	0	0	--
2	6	Collect beads	0.5	8	0	260	--	5	3	0	2	2.5
3	1	Bind	5	8	0	650	--	5	3	0	2	2.5
4	3	Wash 1	2	7	0	600	--	5	3	0	2	2.5
5	4	Wash 2	2	7	6	600	--	5	3	0	2	2.5
6	5	Elution	5	8	0	50	56	5	5	0	2	2.5
7	6	Discard	0.5	5	0	260	--	1	0	0	0	--

2. Related Products

Instrument and Accessories

Product name	Packing Size	Cat.No
TGuide S16 Nucleic Acid Extractor	1 set	OSE-S16-AM
TGuide Smart Magnetic Tip Comb	200 pieces/box	4968939
TGuide Single Sample Tank Bracket	5 pieces/box	4993270

TGuide Smart Reagent Kits

Product name	Packing Size	Cat.No
TGuide Smart Magnetic Plant DNA Kit	48 preps	4993548
TGuide Smart Soil /Stool DNA Kit	48 preps	4993549
TGuide Smart Magnetic Tissue DNA Kit	48 preps	4993547
TGuide Smart Magnetic Plant RNA Kit	48 preps	4993552
TGuide Smart Blood/Cell/Tissue RNA Kit	48 preps	4993551
TGuide Smart Blood Genomic DNA Kit	48 preps	4993703
TGuide Smart Viral DNA/RNA Kit	48 preps	4993702
TGuide Smart Universal DNA Kit	48 preps	4993704